PIP HEALTH FRAUD




PIP PUBLIC CONSULTATION

scenihr

By PIP Action Campaign International Victims Association







Fraudulent PIP Breast Implants: affecting 100,000 women in the EU, the majority of reproductive age, involving CRIMINAL NON COMPLIANT implantable, high risk Category III medical devices of unknown materials, unknown manufacturing processes, unknown risks and documented harm.

Task One Obtain Reliable Data

Task 1: Obtain reliable data on the incidence of implant failure of PIP devices in different countries

SCENIHR concluded that three main factors contribute to implant failure: the physicochemical properties of the implant, the quality of the surgical implantation and the time since implantation. Silicone breast implants may fail, regardless of manufacturer, and the probability of failure increases with time since implantation. Based on clinical studies of PIP implants, the probability of rupture can be estimated to be around 25-30% for PIP implants at 10 years after implantation and with many ruptures occurring or being diagnosed after about 5 years of implantation. Other breast implants from the same calendar time have been found to have an estimated probability of rupture of 2% - 15% after approximately 10 years. Therefore these findings indicate that the manufacturing process of the PIP implants was of inferior quality.

"If you want to encourage surgeons to report adverse incidents with medical devices how does it help if their surgical skills & technique are questioned?"

- Any Fool

VICTIMS & PATIENTS conclude the three main factors contributing to PIP global health scandal:

- 1. Complex Criminal Fraud
- 2. Full scale failure of Regulatory Authorities
- 3. Lack of Reliable Data

SCENIHR is relying on EQUIVALENCE to reach invalid conclusions about **non-compliant**, High Risk Category III Implantable medical devices.


EVIDENCE:

- French Police reports show there were no technically qualified staff at the PIP factory.
- 2. Unknown raw materials
- 3. Unknown manufacturing processes

"There is currently no convincing medical, toxicological or other data to justify removal of intact PIP implants as a precautionary approach."

SCENIHR





Task Two Toxicity of PIP Implants

Task 2: Identify the physicochemical factors that might released due to leakage or influence PIP implant failure in particular the influence of the implant contents routeness toxicological review on the

All silicone based implants contain low molecular weight siloxanes. Batches of PIP devices have been found to contain higher amounts of these siloxanes than those of other manufacturers in particular the cyclic siloxanes D4, D5 and D6. It is however noted that all individuals, regardless of whether or not they have a breast implant, are likely to have significant levels of D4, and D5 in their blood and tissues as these chemicals are widely used in consumer products. Thorough analysis of breast implants has failed to detect any other components of significance. Thus, the questions arise as to whether the higher levels of D4 (octamethylcyclotetrasiloxane), D5 (decamethylpentasiloxane) and D6 in the PIP devices could be the cause of the higher failure rate or could have adverse health consequences for patients when

released due to leakage or rupture. Following a thorough toxicological review on the properties of the two most studied siloxanes, D4 and D5, the conclusion of the SCENIHR is that these compounds are of low acute and chronic toxicity.

VICTIMS & PATIENTS challenge SCENIHR to produce evidence for their highlighted assertions.

EVIDENCE: **Repro-Toxic D4, CAS 556-67-2 Octamethylcyclotetrasiloxane** is currently classified as follows:

- 1. Human health Repro. Cat 3 R62: Possible risk of impaired fertility.
- Human health Hazard class and category: Repr. 2.
 Hazard statement: H361f:
 Suspected of damaging fertility.
- 3. D4 is categorized as an endocrine disruptor (cat 1)

Task Three Adverse Effects on patients


Task 3: Examine whether the adverse effects of a PIP implant failure differ qualitatively and/or qualitatively from the impacts of the failure of breast implants from other manufacturers

The SCENIHR concluded that there is no good evidence that the adverse consequences of a PIP silicone breast implant failure are greater than those resulting from the failure of an implant from another manufacturer. Adverse effects due to free silicone and/or gel-bleed (in the form of siliconomas, lymphadenopathy, lumps etc.) have been reported less uniformly than implant ruptures. The SCENIHR also agreed that there is no indication of a specific association between other effects such as capsular contraction, marked psychological impact on patients and cancer with PIP silicone breast implants. In view of the high rupture rates, many women can expect to experience a ruptured breast implant within their life span, regardless of manufacturer. There is inevitably a higher intra- and postoperative risk associated with the removal of ruptured implants than with intact implants.

VICTIMS & PATIENTS challenges SCENIHR to produce evidence for their highlighted assertions.

EVIDENCE: MHRA (UK) PIP Implant adverse reporting, ANSM (France) published reports, peer-reviewed clinical studies, victims testimonies, PIP Action Health Survey *202.

14. How anxious are you about PIP implants?



"There is currently no convincing medical, toxicological or other data to justify removal of intact PIP implants as a precautionary approach."

SCENIHR





Health Survey *202 Symptoms


Blurred vision	68	36%
Difficulties tolerating bright or fluorescent lights	43	23%
Dry and/or itchy eyes	72	38%
Headaches/ Migraines	108	57%
Poor concentration	119	63%
Memory loss	96	51%
Cognitive loss (difficulty finding the right words)	82	43%
Depression	106	56%
Suicidal thoughts	35	18%
Anxiety	131	69%
Mood swings	114	60%
Anaemia	31	16%
Bleeding gums	54	28%
Tinnitus (ringing in your ears)	46	24%
Pulsatile Tinnitus (hearing your own pulse)	32	17%
Shortness of breath	73	38%
Stiffness or pain in joints	115	61%
Muscle seizures, cramps or spasms	75	39%
Muscle weakness	64	34%
Previously undiagnosed asthma	13	7%

Dry mouth	65	34%
Dry skin	67	35%
Skin rashes	67	35%
Hypersensitivity of skin (can feel like sunburn)	36	19%
Tingling or numbness in hands	96	51%
Swollen joints	36	19%
Excessive sweating	73	38%
Night sweats	98	52%
Extreme tiredness or fatigue	141	74%
Increased bleeding and painful menstrual periods	40	21%
Unexplained absence of menstrual periods	15	8%
Bladder problems	46	24%
Pain in kidneys	32	17%
Bowel Problems	62	33%
Pain in your breasts	128	67%
Itching, tingling or loss of sensation in your breasts	99	52%
Swelling or lumps in your breasts	59	31%
Swelling or lumps in your armpits	60	32%
Hair thinning or hair loss	73	38%
Loss or reduction in sex drive	94	49%
Unexplained lumps and bumps	30	16%
Other	33	17%




"There is currently no convincing medical, toxicological or other data to justify removal of intact PIP implants as a precautionary approach."

SCENIHR







Task Four Knowledge Gaps

Task 4: Knowledge gaps

Registration of all breast implantations on a national or international level could help to collect more evidence for future studies. Continuous surveillance of the different products on the market would enable development and improvements of safe and compliant products as well as be a significant tool in the future scientific research and product monitoring. In order to identify why an implant has caused a significant adverse reaction, it is necessary to analyse the explant involved through a careful selection of explanted devices. The registration of all explantations could be of enormous relevance both from scientific and public health perspective. There is still a need for low cost reliable tests suitable for routine use to

identify implant status (leakage, rupture) in patients.

VICTIMS & PATIENTS challenges SCENIHR to produce evidence for their highlighted assertions.

EVIDENCE:

Criminally NON COMPLIANT PIP Implants have failed to meet the essential requirements referred to in Article 3; Article 8; Article 10; Article 14b, Article 15 (6) and ANNEX I ESSENTIAL REQUIREMENTS I.

GENERAL REQUIREMENTS of the COUNCIL DIRECTIVE 93/42/EEC